Synchronization of Concurrently-Implemented Fluidic Operations in Pin-Constrained Digital Microfluidic Biochips*

Yang Zhao¹, Ryan Sturmer², Krishnendu Chakrabarty¹ and Vamsee K. Pamula²

¹Duke University, Durham, NC 27708, USA

²Advanced Liquid Logic, Inc., Research Triangle Park, NC 27560, USA

Abstract—The implementation of bioassays in pin-constrained biochips may involve pin-actuation conflicts if the concurrently-implemented fluidic operations are not carefully synchronized. We propose a two-phase optimization method to identify and synchronize the fluidic operations that can be executed in parallel. The goal is to implement these fluidic operations without pin-actuation conflict, and minimize the duration of implementing the outcome sequence after the synchronization. The effectiveness of the proposed two-phase optimization method is demonstrated for a representative 3-plex assay performed on a fabricated pin-constrained biochip.

I. INTRODUCTION

Digital microfluidic biochips are emerging as a promising solution for integrating fluid-handling on a chip [1]–[3]. Discrete droplets of nanoliter volumes can be manipulated in a “digital” manner under clock control on a two-dimensional array of electrodes. Compared to traditional bench-top procedures, a microfluidic biochip offers the advantages of low sample and reagent consumption, less likelihood of human error due to automation, high throughput, and high sensitivity.

Recent years have seen a steady increase in the level of integration, system complexity, and applications of digital microfluidic chips [1]. Enzymatic assays [8], DNA sequencing [1], and immunoassays [9] have all been successfully demonstrated on a digital microfluidic chip. Recently, cell-based applications have also been demonstrated on a digital microfluidics platform [3]. These advances in technology serve as a powerful driver for research on computer-aided design (CAD) tools for the design of such chips. Architectural synthesis, physical design, and droplet-routing methods have been developed for the automated design of such chips that can execute laboratory protocols [11]–[16].

In a digital microfluidic biochip, a liquid-based bioassay protocol is implemented through a basis set of operations (such as DISPENSE, TRANSPORT, MIX, SPLIT) by manipulating discrete droplets of nanoliter volumes on a patterned array of electrodes. Each fluidic operation is executed by applying its corresponding pin-actuation sequence to the electrodes. Based on a precomputed bioassay schedule, several fluidic operations are implemented concurrently. If the target biochip is controlled using a direct-addressing scheme, i.e., each electrode is controlled by a separate pin, these fluidic operations can be implemented concurrently without any pin-actuation conflict. However, the direct-addressing scheme leads to a large number of control pins and high fabrication cost [8]. Therefore, several

pin-assignment methods [8], [17]–[19] have been proposed to design pin-constrained biochips.

However, in a pin-constrained biochip design, concurrent implementation of fluidic operations during the same time interval may lead to pin-actuation conflicts, since their corresponding pin-actuation sequences are not completely compatible with each other. Therefore, pin-actuation sequences for the concurrently implemented fluidic operations need to be efficiently synchronized to avoid pin-actuation conflicts. Furthermore, applications such as toxicity screening require rapid time-to-response [7]. Therefore, the length of the outcome sequence (measured in terms of the number of vectors) after synchronization must be minimized.

In this paper, we propose a two-phase optimization method to synchronize the fluidic operations that are executed concurrently. The proposed method is based on fine-grained parallelization of the corresponding pin-actuation sequences, such that pin-actuation conflicts are avoided. The first phase involves the merging of pin-actuation sequences based on clique partitioning in graph theory, and the second phase targets the parallelization of pin-actuation sequences. We propose an integer linear programming (ILP) solution and a heuristic method for the second phase. A multiplexed and concurrent bioassay, namely the n -plex bioassay, is implemented on a fabricated pin-constrained biochip, in order to evaluate the effectiveness of the proposed optimization method.

The rest of the paper is organized as follows. Section II provides an overview of the digital microfluidic platform and pin-constrained biochip design methods. In Section III, we explain the synchronization of concurrently-implemented fluidic operations. In Sections IV–V, we describe the two phases of the proposed optimization method. In Section VI, we demonstrate the effectiveness of the proposed optimization method for the n -plex bioassay. Finally, conclusions are drawn in Section VII.

II. DIGITAL MICROFLUIDIC PLATFORM

In digital microfluidics, droplets of nanoliter volumes, which contain biological samples and reagents, are programmably manipulated on a two-dimensional electrode array [1]. A unit cell in the array includes a pair of electrodes that acts as two parallel plates. The bottom plate contains a patterned array of individually controlled electrodes, and the top plate is coated with a continuous ground electrode. A droplet rests on a hydrophobic surface over an electrode. Coplanar microfluidic devices, i.e., arrays without a top plate, have also been demonstrated [5]. Using the electrowetting phenomenon, droplets can be moved to any location on a two-dimensional array.

Droplets are moved by applying a control voltage to a unit cell adjacent to the droplet and, at the same time, deactivat-

*This work was supported in part by the National Institute of General Medical Sciences of the National Institute of Health (grant # R44GM072155) and the National Science Foundation (grant # CCF-0541055).

Fig. 1. Example of pin-actuation sequences and mapping to an undirected graph.

ing the one just under the droplet. This electronic method of wettability control creates interfacial tension gradients that move the droplets to the activated electrode. Fluid-handling operations such as droplet merging, splitting, mixing, and dispensing can be executed in a similar manner. For example, mixing can be performed by routing two droplets to the same location and then transporting them together [6]. The digital microfluidic platform particularly offers the additional advantage of flexibility, also referred to as reconfigurability, since fluidic operations can be performed anywhere on the array and the droplets need not follow unidirectional pathways. Droplet routes and operation scheduling are pre-programmed into a microcontroller that drives electrodes in the array.

To reduce the number of control pins, several pin-assignment methods [8], [17]–[19] have been proposed to design pin-constrained biochips. In [8], the number of control pins is minimized by using a multi-phase bus for the fluidic pathways. Every n th electrode in an n -phase bus is electrically connected, where n is small number (typically $n = 4$). Thus, only n control pins are needed for a transport bus, irrespective of the number of electrodes that it contains. An alternative design uses row and column addressing, which is referred to as “cross referencing”. An electrode is connected to two pins, corresponding to a row and a column, respectively [17]. A broadcast-addressing-based design technique for pin-constrained and multi-functional biochips is proposed in [19].

III. SYNCHRONIZATION OF CONCURRENTLY-IMPLEMENTED FLUIDIC OPERATIONS IN PIN-CONSTRAINED CHIPS

Based on the schedule derived from architectural synthesis [11], a bioassay is implemented by performing several groups of fluidic operations at different time spans. All the fluidic operations within one group are concurrently implemented by applying their corresponding pin-actuation sequences at the same time. A pin-actuation sequence consists of several pin-actuation vectors that are implemented serially. The time-duration for implementing one vector is one clock cycle. Each bit of a vector represents the status of the corresponding pin in the biochip at this clock cycle. The status of a pin can be either “1” (activate), “0” (deactivate) or “don’t-care”. The “don’t-care” status is represented using the symbol “x”. We use the three values “1”, “0”, and “x” to represent a bit in the pin-activation vector. The *length* of a vector is defined as the number of bits in the vector. The length of each vector in the different sequences is the same; it is equal to the number of control pins in the biochip. Fig. 1(a) shows an example of six pin-actuation sequences, where each sequence consists of six pin-actuation vectors.

Each vector in a pin-actuation sequence contains several don’t-care bits, which can be replaced by “1” or “0”. A *bit-pair* is defined as a pair of two bits that represent the status of the same

pin in two vectors at a given clock cycle. For each bit-pair in two vectors, if either the values of the bit-pair are the same, or the value of one bit in the bit-pair is “x”, we refer to the two vectors as being *compatible*. Two compatible vectors can be made identical by replacing the don’t-care term in each bit-pair with the value (“0” or “1”) of the other bit. If both bits are “x”, we simply map them both to the same logic value. The two identical vectors can then be merged into one pin-actuation vector.

Consider a group of pin-actuation sequences corresponding to a group of concurrently-implemented fluidic operations. For each time-step, if all the vectors are pair-wise compatible, we refer to the sequences in the group as being *compatible*. Mutually-compatible sequences can be implemented concurrently, without any conflict in pin actuation. Therefore, these mutually-compatible sequences in the group can be made identical and merged into one outcome sequence by serially merging all the vectors at the same time-step.

If the pin-actuation sequences in a group are not mutually compatible, the merging of these sequences may lead to erroneous assay outcomes. Therefore, we present a two-phase method for fine-grained merging and parallelization of such sequences. In the first phase, we partition these sequences into groups, such that all sequences in a group are pairwise-compatible and can be merged into a single outcome sequence. However, these merged outcome sequences are not mutually compatible. In the second phase, by stalling some vectors in the outcome sequences of Phase 1 for several clock cycles, we can further parallelize these outcome sequences. A vector is *stalled* for k clock cycles if it is repeated k times, thereby delaying the successor vectors in the sequence. An optimization method is used to minimize the number of vectors in the sequence resulting from Phase 2.

IV. MERGING OF SEQUENCES (PHASE 1)

The problem of finding the minimum number of groups can be easily mapped to the clique-partitioning problem from graph theory [4]. We use the example in Fig. 1(a) to illustrate this mapping. Based on the actuation-sequence table, we construct an undirected graph, referred to as *pin-actuation graph*; see Fig. 1(b). Each node in the graph represents a pin-actuation sequence. There is an edge between a pair of nodes if their corresponding pin-actuation sequences are compatible.

Clique partitioning refers to the problem of dividing a set of nodes into non-overlapping subsets, such that the subgraph constructed by each subset is a complete subgraph. A *minimal clique partition* is one that covers the nodes in the graph with a minimum number of disjoint complete subgraphs. The minimum number of groups of compatible sequences is equivalent to the size of a minimal clique-partition. A minimal clique partition for the example in Fig. 1 is given by $\{1,3,5\}$, $\{2,4,6\}$. Although the clique partitioning problem is known to be NP-hard [4], heuristics can be used to solve it in an efficient manner.

Next we merge all the sequences in a clique to an outcome sequence. For example, for the group $\{1,3,5\}$, the outcome sequence “011, 111, 000, 011, 100, 100” is compatible with all the individual sequences. Therefore, we can concurrently execute the operations corresponding to sequences 1, 3, and 5. The group $\{2,4,6\}$ can be merged into an outcome sequence “110, 011, 100, 011, 011, 100”.

However, the merged sequences are not mutually compatible. Therefore, they cannot be performed concurrently. We next parallelize these non-mutually-compatible sequences, and minimize the number of vectors in the resulting outcome sequence.

V. PARALLELIZATION OF SEQUENCES (PHASE 2)

In order to parallelize incompatible pin-actuation sequences, we stall some vectors in the sequences for an appropriate number of clock cycles, so as to make all vectors at the same time-step pairwise-compatible. After such fine-grained parallelization, the length of a sequence may be increased since some of its vectors may be stalled for several clock cycles. However, the implementation order of the vectors in each sequence remains the same. In the extended sequences, the vectors at the same time-step are pair-wise compatible. Therefore, these extended sequences are mutually-compatible and they can be merged to form an outcome sequence. Note that at a specific clock cycle, a vector at the previous clock cycle from one sequence can stall only when it is mutually compatible with vectors at current clock cycle from other sequences. Without the above constraint, i.e., vectors at a specific clock cycle are not mutually compatible, the concurrent implementation of these vectors will lead to pin-actuation conflicts.

However, stalling a vector, which was intended for only one clock cycle in the assay protocol, for several clock cycles will increase the completion time of its corresponding operation; therefore, it increases the completion time for a group of operations that are executed concurrently. Therefore, the length (number of vectors) of the outcome sequence after parallelization must be minimized.

A. Parallelization based on an ILP model

Assume that N concurrently-implemented pin-actuation sequences need to be parallelized, and let the length of sequence k be L_k . Sequence k includes a series of L_k vectors, namely $V_{i_k} (i_k = 1, 2, \dots, L_k)$. A set of vectors is denoted by $\{V_{i_1}, V_{i_2}, \dots, V_{i_N}\}$, where V_{i_k} is the i_k th vector from sequence k .

If vector V_{i_k} from sequence k is not compatible with another vector V_{i_l} from sequence l , the vectors in a vector set that includes V_{i_k} and V_{i_l} are not mutually compatible. A vector set is called a *compatible vector set* only if all the vectors within it are mutually compatible. Let $c_{i_1 i_2 \dots i_N}$ be a constant defined as follows:

$$c_{i_1 i_2 \dots i_N} = \begin{cases} 1 & \text{if the vectors in the set } \{V_{i_1}, V_{i_2}, \dots, V_{i_N}\} \\ & \text{are all mutually compatible} \\ 0 & \text{otherwise} \end{cases}$$

For a compatible vector-set, we determine if all vectors in it can be merged to form a vector of the outcome sequence. The parameter $c_{i_1 i_2 \dots i_N}$ can be pre-calculated from the given vector sequences.

Let $x_{i_1 i_2 \dots i_N}$ be a binary variable defined as follows:

$$x_{i_1 i_2 \dots i_N} = \begin{cases} 1 & \text{if the vectors in the set} \\ & \{V_{i_1}, V_{i_2}, \dots, V_{i_N}\} \text{ are merged} \\ 0 & \text{otherwise} \end{cases}$$

Our goal is to minimize the number of vectors in the outcome sequence after parallelization. We achieve this goal by minimizing

the number of merged vector sets, where all vectors in a set are mutually compatible. The objective function for the ILP model can therefore be stated as follows:

$$\text{Minimize } \mathcal{F} = \sum_{i_1=1}^{L_1} \sum_{i_2=1}^{L_2} \dots \sum_{i_N=1}^{L_N} c_{i_1 i_2 \dots i_N} x_{i_1 i_2 \dots i_N}.$$

The corresponding operations that are executed concurrently should start and finish at the same time, since the droplets involved in these operations should be maintained at their initial or final locations by avoiding the “floating” status of pins. If an operation does not start while other concurrently-executed operations start, its pin-actuation is not applied and the corresponding control pins are in “floating” status. In this situation, the droplets involved in this operation will not be held on the initial positions and may mix inadvertently with other droplets.

Therefore, all vectors in sets $\{V_1, V_1, \dots, V_1\}$ and $\{V_{L_1}, V_{L_2}, \dots, V_{L_N}\}$ must be mutually compatible, i.e., the conditions $c_{11\dots 1} = 1$ and $c_{L_1 L_2 \dots L_N} = 1$ must be satisfied. We add constraints to ensure that all vectors in sets $\{V_1, V_1, \dots, V_1\}$ and $\{V_{L_1}, V_{L_2}, \dots, V_{L_N}\}$ are merged. The constraints can be represented by the following two equations: (i) $c_{11\dots 1} x_{11\dots 1} = 1$; (ii) $c_{L_1 L_2 \dots L_N} x_{L_1 L_2 \dots L_N} = 1$.

To perform the N pin-actuation sequences synchronously, each vector in a sequence (e.g., vector V_{i_k} in k th sequence) must be executed for at least one clock cycle during the implementation of the outcome sequence after parallelization. Therefore, vector V_{i_k} must be in at least one compatible vector-set, which must be merged to form a vector in the outcome sequence. This constraint can be written as: $\forall i_k (1 \leq i_k \leq L_k, 1 \leq k \leq N)$,

$$\sum_{i_1=1}^{L_1} \dots \sum_{i_{k-1}=1}^{L_{k-1}} \sum_{i_{k+1}=1}^{L_{k+1}} \dots \sum_{i_N=1}^{L_N} c_{i_1 \dots i_N} x_{i_1 \dots i_N} \geq 1.$$

Consider a compatible vector-set $vs_1: \{V_{i_1}, \dots, V_{i_k}, \dots, V_{i_l}, \dots, V_{i_N}\}$, where $1 \leq i_k \leq L_k, 1 \leq i_l \leq L_l$ and $1 \leq k, l \leq N$, which has already been merged. That is to say, $x_{i_1 \dots i_k \dots i_l \dots i_N} = 1$. If two vectors V_{j_k} and V_{j_l} in another compatible vector-set $vs_2: \{V_{j_1}, \dots, V_{j_k}, \dots, V_{j_l}, \dots, V_{j_N}\}$ satisfy the relationship “ $1 \leq j_k < i_k$ ” and “ $i_l < j_l \leq L_l$ ”, vs_2 must not be merged. In other words, $x_{j_1 \dots j_k \dots j_l \dots j_N}$ must be equal to 0.

The violation of the above constraint will lead to timing problems when we implement N sequences concurrently. For instance, suppose the compatible vector-set vs_2 can be merged, i.e., $x_{j_1 \dots j_k \dots j_l \dots j_N} = 1$. Under this assumption, two vectors obtained by merging both vector sets vs_1 and vs_2 will be implemented in the outcome sequence serially. However, on one hand, if the vector obtained by merging vs_1 is implemented before the vector obtained by merging vs_2 , the prerequisite “ $1 \leq j_k < i_k$ ” is violated. On the other hand, implementing the vector obtained by merging vs_2 before the vector obtained by merging vs_1 will violate the prerequisite “ $i_l < j_l \leq L_l$ ”. Therefore, only one compatible vector-set, either vs_1 or vs_2 , can be merged. We formulate this constraint as follows:

$$c_{i_1 \dots i_k \dots i_l \dots i_N} x_{i_1 \dots i_k \dots i_l \dots i_N} + c_{j_1 \dots j_k \dots j_l \dots j_N} x_{j_1 \dots j_k \dots j_l \dots j_N} \leq 1.$$

The complete ILP model is as follows:

Minimize $\mathcal{F}$, subject to

$$(1) \ c_{11\dots 1} x_{11\dots 1} = 1.$$

TABLE I
EXAMPLE OF PIN-ACTUATION SEQUENCES.

Sequence. no.	1	2	3
Actuation sequence	1xx1	x11x	1x1x
	1x10	x01x	xx11
	0x11	1x1x	x01x
	x0x1	0xxx	x01x
	00xx		

$$(2) c_{L_1 L_2 \dots L_N} x_{L_1 L_2 \dots L_N} = 1.$$

$$(3) \forall i_k (1 \leq i_k \leq L_k, 1 \leq k \leq N),$$

$$\sum_{i_1=1}^{L_1} \dots \sum_{i_{k-1}=1}^{L_{k-1}} \sum_{i_{k+1}=1}^{L_{k+1}} \dots \sum_{i_N=1}^{L_N} c_{i_1 \dots i_N} x_{i_1 \dots i_N} \geq 1.$$

$$(4) \text{ if } 1 \leq j_k < i_k \text{ and } i_l < j_l \leq L_l,$$

$$c_{i_1 \dots i_k \dots i_l \dots i_N} x_{i_1 \dots i_k \dots i_l \dots i_N} + c_{j_1 \dots j_k \dots j_l \dots j_N} x_{j_1 \dots j_k \dots j_l \dots j_N} \leq 1.$$

We use an example in Table I to illustrate the parallelization of pin-actuation sequences for $N = 3$. There are $5 \times 4 \times 4$ vector sets, i.e., 80 vector sets. We calculate $c_{i_1 i_2 i_3}$ ($1 \leq i_1 \leq 5, 1 \leq i_2 \leq 4, 1 \leq i_3 \leq 4$) in advance before constructing the ILP model. For example, all vectors “1xx1”, “x11x”, and “1x1x” in the vector set $\{V_1, V_1, V_1\}$ are mutually compatible, therefore $c_{111} = 1$. The binary variable $x_{i_1 i_2 i_3}$ is 1 if vectors V_{i_1}, V_{i_2} , and V_{i_3} are merged; otherwise it is 0. The objective function is as follows:

$$\text{Minimize } \mathcal{F} = \sum_{i_1=1}^5 \sum_{i_2=1}^4 \sum_{i_3=1}^4 c_{i_1 i_2 i_3} x_{i_1 i_2 i_3}.$$

The constraints that all vectors in sets $\{V_1, V_1, V_1\}$ and $\{V_5, V_4, V_4\}$ must be mutually compatible are: (i) $c_{111} x_{111} = 1$; (ii) $c_{544} x_{544} = 1$. The constraint that each vector in a sequence (e.g., the 2nd vector in the 2nd sequence) must be in at least one merged vector-set must also be satisfied. For example, for the 2nd vector in Seq. 2, we have $\sum_{i_1=1}^5 \sum_{i_3=1}^4 c_{i_1 2 i_3} x_{i_1 2 i_3} \geq 1$.

We next consider the constraints for avoiding timing problems. For example, the compatible vector set $\{V_{i_1}, V_{i_2}, V_{i_3}\} = \{V_4, V_2, V_1\}$, and compatible vector set $\{V_{j_1}, V_{j_2}, V_{j_3}\} = \{V_1, V_2, V_3\}$, satisfy the relationship “ $j_1 < i_1$ ” and “ $i_3 < j_3$ ”. Suppose the two compatible vector-sets can be merged at the same time. Two vectors obtained by merging both vector sets will be implemented in the outcome sequence serially. If the vector obtained by merging $\{V_4, V_2, V_1\}$ is implemented before the vector obtained by merging $\{V_1, V_2, V_3\}$, the 4th vector from Seq. 1 is implemented before the 1st vector from Seq. 1, which violates the order of vectors in Seq. 1. On the other hand, if we implement the vector obtained by merging $\{V_1, V_2, V_3\}$ before the vector obtained by merging $\{V_4, V_2, V_1\}$, the 3rd vector from Seq. 3 is implemented before the 1st vector from Seq. 3, which violates the order of vectors in Seq. 3. Therefore, only one compatible vector-set, either $\{V_1, V_2, V_3\}$ or $\{V_4, V_2, V_1\}$, can be merged. This constraint is written as follows: $c_{123} x_{123} + c_{421} x_{421} \leq 1$.

The results obtained by solving the ILP model are shown in Table II. If $x_{i_1 i_2 i_3} = 1$, a group of mutually compatible vectors V_{i_1}, V_{i_2} , and V_{i_3} are merged into a vector of the outcome sequence (Outcome Seq.). Note that some vectors are stalled for several clock cycles to facilitate parallelization. For example, the first vectors in Seq. 2 and Seq. 3 are stalled for two clock cycles

TABLE II
RESULT OBTAINED FROM ILP MODEL.

Clock cycle	$x_{i_1 i_2 i_3}$	Sequence no.			
		Seq. 1	Seq. 2	Seq. 3	Outcome Seq.
1	$x_{111} = 1$	1xx1	x11x	1x1x	1111
2	$x_{211} = 1$	1x10	x11x	1x1x	1110
3	$x_{322} = 1$	0x11	x01x	xx11	0011
4	$x_{433} = 1$	x0x1	1x1x	x01x	1011
5	$x_{544} = 1$	00xx	0xxx	x01x	001x

- 1: Check whether vectors in set $\{V_1, V_1, \dots, V_1\}$ are mutually compatible and merge them together to form the first vector in outcome sequence;
- 2: **for** $j = 1$ **to** *ROUND* **do**
- 3: Reset all the pointers to second vectors of sequences;
- 4: **while** at least one vector has not been parallelized **do**
- 5: Check the location of each pointer, e.g., pointer of the k th pin-actuation sequence ($1 \leq k \leq N$) locates at V_{i_k+1} ;
- 6: Randomly select a compatible vector-set from vector-sets $\{V_{i_1} \text{ or } V_{i_1+1}, V_{i_2} \text{ or } V_{i_2+1}, \dots, V_{i_k} \text{ or } V_{i_k+1}, \dots, V_{i_N} \text{ or } V_{i_N+1}\}$ except the vector-set $\{V_{i_1}, V_{i_2}, \dots, V_{i_k}, \dots, V_{i_N}\}$;
- 7: **if** a compatible vector-set exists **then**
- 8: Merge all the vectors within it to form a vector in the outcome sequence; Adjust the pointer for each sequence;
- 9: **else**
- 10: break;
- 11: **end if**
- 12: **end while**
- 13: **if** finish parallelization of N pin-actuation sequences **then** break;
- 14: **end if**
- 15: **end for**
- 16: Output the result of parallelization: success or failure.

Fig. 2. Pseudocode to parallelize N pin-actuation sequences.

each. The experiments were performed on a 2.0 GHz INTEL Core2 Dual processor, with 1 GB of memory. The CPU time needed for solving the above ILP model is only 0.1s for this example.

Even though the above ILP model is useful for problem understanding and for deriving lower bounds on sequence length via LP-relaxation, it suffers from the drawback that it may be too complex for problem instances involving more than a few pin-actuation sequences or vectors. Therefore, we develop a heuristic method to parallelize N sequences within reasonable CPU time.

B. Parallelization based on a heuristic method

For each sequence, we assign a pointer to the vector whose successor vectors (including itself) have not been parallelized, while all its predecessor vectors have been parallelized with vectors from other sequences into vectors of the outcome sequence.

At first, each pointer is at the first vector of each sequence. We check whether the first vectors in all the sequences, i.e., the vector-set $\{V_1, V_1, \dots, V_1\}$ are mutually compatible and merge them together to form the first vector in outcome sequence. Next, all pointers independently move to the 2nd vectors of all the sequences, indicating that the 2nd vectors and their successor vectors in the sequences have not been parallelized.

Assume that for the k th pin-actuation sequence ($1 \leq k \leq N$), vectors $V_1, V_2, \dots, V_{i_k}$ have been parallelized with vectors from other sequences into vectors of the outcome sequence, while vectors $V_{i_k+1}, V_{i_k+2}, \dots, V_{L_k}$ have not yet been parallelized. For all N sequences, all vectors in vector-set $\{V_{i_1}, V_{i_2}, \dots, V_{i_k}, \dots, V_{i_N}\}$ are mutually compatible and have been merged to form a vector in the outcome sequence. Currently the pointer of the k th

Fig. 3. Fabricated biochip used for the n -plex assay experiment.

pin-actuation sequence ($1 \leq k \leq N$) is at V_{i_k+1} . We attempt to find a compatible vector-set within vector-sets $\{V_{i_1}$ or V_{i_1+1} , V_{i_2} or V_{i_2+1} , $\dots$, V_{i_k} or V_{i_k+1} , $\dots$, V_{i_N} or $V_{i_N+1}\}$, except the vector-set $\{V_{i_1}, V_{i_2}, \dots, V_{i_k}, \dots, V_{i_N}\}$. If a compatible vector-set exists, e.g., all the vectors in vector-set $\{V_{i_1+1}, V_{i_2+1}, \dots, V_{i_k+1}, \dots, V_{i_N+1}\}$ are mutually compatible, we can merge these vectors together to form a vector in the outcome sequence. If multiple compatible vector-sets exist, we randomly select one set and merge all the vectors within it. For the k th pin-actuation sequence ($1 \leq k \leq N$), if V_{i_k+1} is in the selected vector-set, it is merged with vectors from other sequences to form a vector in the outcome sequence, and the k th pointer is moved to the immediate successor vector V_{i_k+2} . However, if V_{i_k} is with the selected vector-set, it is merged with vectors from other sequences. In this case, the k th pointer remains at the vector V_{i_k+1} , since V_{i_k+1} has not been parallelized. The above process is repeated until all the pointers reach the end of the sequences.

However, if no compatible vector-set exists, it means that the parallelization of N pin-actuation sequences cannot proceed any more. In this case, we reset all the pointers to the second vectors of sequences, thereby restarting the parallelization process. After multiple rounds, if there is still no compatible vector-set at a specific clock cycle, we conclude that the heuristic method has failed to parallelize the sequences. This aspect of the heuristic method is a drawback compared to the ILP model.

The pseudocode for the heuristic method is shown in Fig. 2, where *ROUND* is a user defined parameter. The worst-case time complexity of the heuristic method for N pin-actuation sequences with maximum sequence length L_{max} is $O(ROUND \times N \times L_{max})$.

VI. EXPERIMENTAL RESULTS

The n -plex bioassay is a typical example of a multiplexed and concurrent bioassay. In this assay, a sample is analyzed for n different analytes. Sample droplets are mixed with n different reagents, and the mixed product droplets are routed to a detection site. The signal transduction for quantifying the n droplets is based on an optical signal from each droplet. An optical detector such as a photodiode can be used to analyze the product droplets of the n -plex bioassay.

As an example of an n -plex assay, we demonstrated a 3-plex assay for the diagnosis of acute myocardial infarction (AMI) [10] on a fabricated digital microfluidic platform; see Fig. 3. The fabricated chip is a pin-constrained design that consists of over 1000 electrodes and 64 control pins [10]. The platform consists of three regions: dispensing region, reaction region, and detection

region. Note that the electrodes in different regions are connected to separate sets of control pins. In the dispensing region, there are 20 reservoirs for high-throughput processing, with 8 reservoirs for sample solutions and 12 reservoirs for reagent solutions. The sample (reagent) reservoirs are accessed using vertical (horizontal) transport buses. Each reagent reservoir intersects with these vertical transport buses using a horizontal transport bus. In the reaction region located in the middle of the platform, there are 12 horizontally-placed reactors where the reagent and sample droplets are mixed. Each reactor is connected to a horizontal transport bus from the dispensing region. The droplets are then routed through a multi-phase transporter in the detection region.

For the 3-plex assay, an example of human physiological fluids, i.e., serum, is sampled and dispensed to the digital microfluidic biochip. Three assays, namely troponin-I, myoglobin, and creatine kinase-MB (CK-MB) measurements are performed on the physiological fluid. The 3-plex assay includes three stages: dispensing and routing, reaction, and detection, which are performed in the dispensing region, reaction region, and detection region, respectively.

The 3-plex assay is implemented by serially performing groups of fluidic operations at different time spans. All the fluidic operations within one group are concurrently implemented by applying their corresponding pin-actuation sequences at the same time. We synchronize these operations by merging and parallelizing their corresponding pin-actuation sequences, in order to avoid pin-actuation conflicts. For example, during a specific duration in the detection stage, seven operations, namely dispensing a substrate droplet (denoted as *SB*) from the substrate reservoir, dispensing a wash droplet (denoted as *WA*) from the wash reservoir, transporting two product droplets D_1 and D_2 , as well as transporting three wash droplets W_1 , W_2 , W_3 along the transport bus, are implemented concurrently. The details about their corresponding pin-actuation sequences are described in Table III.

We next synchronize the operations in Table III by merging and parallelizing their corresponding pin-actuation sequences. We construct a pin-actuation graph based on the sequences in Table III; see Fig. 4(a). Note that a group of nodes corresponding to sequences 3-7 are pair-wise compatible, since each of them is the sequence corresponding to the operation that transports a droplet with 8-electrode spacing. A minimal clique partition for the example in Fig. 4(a) is given by $\{1\}$, $\{2\}$, $\{3, 4, 5, 6, 7\}$. We merge all the sequences in the third partition. The outcome sequence obtained by merging sequences in the partition $\{3, 4, 5, 6, 7\}$ is the same as Seq. 3.

We next parallelize the three sequences, namely Seq. 1-3, using both the ILP model and the heuristic method. The CPU times for solving this ILP model is 26.12 s. We construct a table based on the solution obtained using the ILP model, as shown in Fig. 4(b). If $x_{i_1 i_2 i_3} = 1$, a group of vectors V_{i_1} , V_{i_2} , and V_{i_3} are pair-wise compatible at the same time-step, therefore they are listed in the same row. Note that some vectors are stalled for several clock cycles to facilitate the parallelization. The vectors in each row are merged to form a vector of the outcome sequence. The outcome sequence is generated by merging all the vectors in the same row into an outcome vector serially from clock cycle 1 to 9. The length of the outcome sequence is minimized to 9 clock cycles,

TABLE III
PIN-ACTUATION SEQUENCES CORRESPONDING TO OPERATIONS IN DETECTION STAGE.

Clock cycle	Sequence. no. (operation)						
	Seq. 1 (W_A)	Seq. 2 (S_B)	Seq. 3 (D_1)	Seq. 4 (D_2)	Seq. 5 (W_1)	Seq. 6 (W_2)	Seq. 7 (W_3)
1	V_{11}	V_{21}	V_{31}	V_{41}	V_{51}	V_{61}	V_{71}
2	V_{12}	V_{22}	V_{32}	V_{42}	V_{52}	V_{62}	V_{72}
3	V_{13}	V_{23}	V_{33}	V_{43}	V_{53}	V_{63}	V_{73}
4	V_{14}	V_{24}	V_{34}	V_{44}	V_{54}	V_{64}	V_{74}
5	V_{15}	V_{25}	V_{35}	V_{45}	V_{55}	V_{65}	V_{75}
6	V_{16}	V_{26}	V_{36}	V_{46}	V_{56}	V_{66}	V_{76}
7		V_{27}	V_{37}	V_{47}	V_{57}	V_{67}	V_{77}
8			V_{38}	V_{48}	V_{58}	V_{68}	V_{78}
9			V_{39}	V_{49}	V_{59}	V_{69}	V_{79}

Fig. 4. Merging and parallelization of pin-actuation sequences in Table III: (a) pin-actuation graph; (b) result of ILP model.

Fig. 5. Comparison of the completion time of the proposed optimization method with the baseline schedule for the 3-plex assay.

equal to the length of Seq. 3.

We next use the heuristic method to parallelize Seq. 1-3. The CPU time is only 0.37 s, which is 1.5% of the CPU time for ILP. The length of the outcome sequence is 10 clock cycles, one clock cycle larger than that for the ILP model.

We next compare the completion time of the 3-plex bioassay for the ILP model, the heuristic method, and a baseline method; see Fig. 5. For the baseline method, we do not consider the merging and parallelization of a group of concurrently-implemented fluidic operations at each time span. Since the pin-actuation sequences of those operations are not mutually compatible, they have to be implemented serially. Our results show a significant reduction (nearly one-third) in the time-to-result for the 3-plex bioassay.

VII. CONCLUSIONS

We have presented a two-phase optimization method to synchronize the fluidic operations that are executed concurrently

in pin-constrained biochips, in order to avoid pin-actuation conflicts. The optimization method merges the corresponding pin-actuation sequences based on clique partitioning in graphs in the first phase, and parallelizes the non-mutually-compatible pin-actuation sequences in the second phase. The optimization method has also minimized the length of the outcome sequence after the synchronization. A 3-plex bioassay has been implemented on a fabricated pin-constrained biochip to demonstrate the effectiveness of the proposed optimization method.

REFERENCES

- [1] R. B. Fair et al., "Chemical and biological applications of digital-microfluidic devices", *IEEE Design & Test of Computers*, vol. 24, 2007.
- [2] J. Gong et al., "Portable digital microfluidics platform with active but disposable lab-on-chip," *Proc. IEEE MEMS*, pp. 355-358, 2004.
- [3] I. B.-Nad et al., "Digital microfluidics for cell-based assays", *Lab on a Chip*, vol. 8, pp. 519-526, 2008.
- [4] J. L. Gross and J. Yellen, *Graph Theory and Its Applications*, CRC Press, FL, 1999.
- [5] P. Paik et al., "Coplanar digital microfluidics using standard printed circuit board processes," *MicroTAS*, 2005.
- [6] P. Paik et al., "Electrowetting-based droplet mixers for microfluidic systems", *Lab on a Chip*, vol. 3, pp. 28-33, 2003.
- [7] K. Gernaey et al., "Fast and sensitive acute toxicity detection with an enrichment nitrifying culture", *Water Environment Research*, vol. 69, pp. 1163-1169, 1997.
- [8] V. Srinivasan et al., "An integrated digital microfluidic lab-on-a-chip for clinical diagnostics on human physiological fluids", *Lab on a Chip*, vol. 4, pp. 310-315, 2004.
- [9] R. Sista et al., "Heterogeneous immunoassays using magnetic beads on a digital microfluidic platform", *Lab on a Chip*, vol. 8, pp. 2188-2196, 2008.
- [10] R. Sista et al., "Development of a digital microfluidic platform for point of care testing", *Lab on a Chip*, vol. 8, pp. 2091-2104, 2008.
- [11] F. Su and K. Chakrabarty, "High-level synthesis of digital microfluidic biochips", *ACM J. Emerging Tech. Comp. Sys.*, vol. 3, 2008.
- [12] F. Su and K. Chakrabarty, "Unified high-level synthesis and module placement for defect-tolerant microfluidic biochips", *Proc. DAC*, 2005.
- [13] F. Su et al., "Droplet routing in the synthesis of digital microfluidic biochips", *Proc. DATE*, pp. 323-328, 2006.
- [14] P.-H. Yuh et al., "Placement of defect-tolerant digital microfluidic biochips using the T-tree formulation", *ACM J. Emerging Tech. Computing Sys.*, vol. 3, pp. 13.1-13.32, 2007.
- [15] P.-H. Yuh et al., "BioRoute: A network flow based routing algorithm for digital microfluidic biochips", *Proc. ICCAD*, pp. 752-757, 2007.
- [16] M. Cho and D. Z. Pan, "A high-performance droplet router for digital microfluidic biochips", *Int. Symp. Phys. Design*, pp. 200-206, 2008.
- [17] S.-K. Fan et al., "Manipulation of multiple droplets on $N \times M$ grid by cross-reference EWOD driving scheme and pressure-contact packaging", *Proc. Int. Conf. MEMS*, pp. 694-697, 2003.
- [18] T. Xu et al., "Automated design of pin-constrained digital microfluidic biochips under droplet-interference constraints", *ACM J. Emerging Tech. in Computing Sys.*, vol. 3, article 14, 2007.
- [19] T. Xu and K. Chakrabarty, "Broadcast electrode-addressing for pin-constrained multi-functional digital microfluidic biochips", *Proc. DAC*, pp. 173-178, 2008.